

SCALP VEIN INFUSION SETS

YOUR CHOICE DEFINES YOUR SAFETY NEED



Product Range

Needle Gauge	Colour Code	Needle Length (mm) & (inch)		Needle Dia. (mm)
18G	Pink	19	3/4"	1.20
19G	Cream	19	3/4"	1.10
20G	Yellow	19	3/4"	0.90
21G	Green	19	3/4"	0.80
22G	Black	19	3/4"	0.70
23G	Light Blue	19	3/4"	0.60
24G	Medium purple	19	3/4"	0.55
25G	Orange	19	3/4"	0.50
26G	Brown	19	3/4"	0.45


Components

1. Luer Adaptor - Luer adaptors are transparent which allow visual monitoring of the flow of the medicaments. Luer adaptors are having a 6% luer taper as per ISO-594 which provides a leak free mating with an infusion set or a syringe of any make and size.

LUER CONNECTOR: Presence of soft PVC luer connector which provides additional secure fitment & prevents an accidental leakage.

3. WINGS - The soft wings conform well to the skin contours and make immobilization simple.

4. LINK TUBE - The link tube is translucent for easy observation of the contents.



5 NEEDLE - The needle is thin walled which allows a larger internal diameter and is ideal for rapid infusion and dosage with high viscosity medications. The needle has a short bevel three facet point to minimize the chances of cross perforation of vein during insertion. Needles are siliconised to reduce friction & pain (for smooth penetration and withdrawal).

PACKING

Individually packed in transparent thermoformed peelable blister pack.

	UNIT PACK	SHELF CARTON	SHIPPING CORRUGATED BOX
SCALP VAN	Two blister of SCALP VAN S.V. sets from a strip each unit is easily separable from the strip along with perforated slitting	50 units of SCALP VAN S.V. sets (25 strips of two units each are packed in a printed duplex board box).	40 shelf cartons are packed in a 5 ply corrugated carton pack Size: 565x465x320 Gross Wt.: 10.4 kg.(approx.)

WARNING

- Do not use if the pack is previously opened or damaged.
- Do not attempt to re-sterilize and use.
- Infusion sets should not be recycled and should be disposed off preferably in sharps container.

Precautions

Before use check the integrity of the sealing of the pack. Tape the wings securely after insertion.